

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
Document Name	Document Type
	Page 1 of 29

# **HC2-LDP**

## **Low Dewpoint Probe**

### **Instruction Manual**



<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
Document Name	Document Type

Page 2 of 29

## Table of Contents

<b>Table of Contents .....</b>	<b>2</b>
<b>1      Overview .....</b>	<b>3</b>
1.1     Applicability.....	4
1.2     Background Information on the Principle of Measurement.....	4
<b>2      Mechanical Dimensions.....</b>	<b>7</b>
2.1     Probe.....	7
2.2     Measurement Chamber.....	8
<b>3      General.....</b>	<b>9</b>
3.1     Power Supply .....	9
3.2     Measured Parameters .....	9
3.3     UART Interface.....	9
3.4     HW4 Software Compatibility .....	10
3.5     Sensor Filter .....	10
<b>4      User-Specific Settings and Functions.....</b>	<b>11</b>
4.1     General AirChip4000 Functions.....	11
4.2     Factory Defaults .....	14
<b>5      Mechanical Installation.....</b>	<b>15</b>
5.1     Choice of Measurement Point .....	15
5.2     Installation of the Probe .....	17
5.3     Pin Configuration of Plug Connector .....	18
<b>6      Maintenance.....</b>	<b>19</b>
6.1     Service Cable .....	19
6.2     Calibration .....	19
6.3     Probe.....	20
<b>7      Firmware Update .....</b>	<b>21</b>
<b>8      Technical Data .....</b>	<b>22</b>
8.1     Technical data, measurement chamber .....	24
<b>9      Accessories .....</b>	<b>26</b>
<b>10     Additional Documents .....</b>	<b>28</b>
<b>11     Document Approval .....</b>	<b>29</b>

<b>E-M-HC2-LDP-V1_06</b>		
Document Code		
<b>HC2-LDP Manual</b>		<b>Instruction manual</b>
		Document Type
Document Name		Page <b>3</b> of 29

## 1 Overview

The ROTRONIC low-dewpoint probe is designed to measure trace moisture in the range of -70 .. 85 °C (-94 .. 185 °F)  $T_d$ . The dew point probes are based on a capacitive trace moisture sensor, a Pt1000 temperature sensor, and the high-precision AirChip4000 measurement electronics for evaluation and processing of the sensor data. The probe is designed for dewpoint/frostpoint measurements in compressed-air systems. The robust stainless-steel housing and the stainless-steel sinter filter make the probe suitable for industrial applications. The output signals from the low-dewpoint probe are exclusively digital. Suitable devices for making the dewpoint/frostpoint and temperature measurements available as analog signals are the ROTRONIC measurement converters in the HF5, HF8 and PF4 series, as well as the ROTRONIC HP22 and HP23 hand-held meters. Alternatively, the probe can be connected directly to a PC USB port, and read out using MS Windows HW4 software.

The low-dewpoint probe has the following features:

- Very high precision measurement of dewpoint, frostpoint and temperature
- Pressure-resistant to 100 bar
- Absolute repeatability guaranteed
- The latest AirChip4000 measurement technology
- HYGROMER® LDP-1 sensor
- Software-based sensor alarm function
- Adjustment by customer possible
- 1-point temperature calibration / adjustment
- Multi-point dew point calibration / adjustment

The units in the LDP series can be kept up to date with regard to improvements in functionality through simple update of the firmware.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page 4 of 29

## 1.1 Applicability

This manual applies to LDP series probes with firmware version V1.5. The low-order digit of the firmware version stands for minor changes, e.g. correction of errors which do not influence the main functionality of the device.

## 1.2 Background Information on the Principle of Measurement

The presence of liquid water is undesirable in many industrial gases (e.g. compressed air). It can lead to rust in the piping and functional elements, rupture of lubricant films, electrochemical formation of elements, microbial contamination, and icing, with the associated damage to the system. If the dewpoint of the process gas is known, a suitable drying process can be used to lower its (pressure) dewpoint to a point at which the formation of liquid water (condensation) is avoided. By and large, it can be said that the lower the temperature and humidity of a gas is, the lower its pressure dewpoint will be. Determining this kind of low dewpoint, especially at high gas pressure, makes considerable demands on the pressure resistance, chemical resistance, and resolution of the humidity probe, and on the processor power necessary for the electronics to convert temperature and gas humidity to the associated low-dewpoint temperature.

That is why the latest AirChip4000 technology is used in the ROTRONIC low-dewpoint probe. It is based on an enhancement and power boost to the market-proven, high-precision AirChip3000 technology. Humidity measurement is effected by the new, capacitive HYGROMER® LDP-1 sensor. Together with the Pt1000 temperature sensor, it allows the dewpoint or frostpoint to be determined precisely. The AirChip4000 technology can ascertain any of over 8,000 individual points, and guarantees a high degree of repeatability. The increased performance is vital for dewpoint applications, because the slightest variations in the environmental conditions of the delicate trace moisture climate are very important for the results of measurement.

The firmware of the low-dewpoint probe also has an integrated low-pass filter to effectively reduce signal-noise interference. The result is a low-dewpoint measurement precision of  $\pm 2 \text{ K } T_d$  in the range -50 .. 20 °C  $T_d$ .

Familiar functions of the AirChip3000 technology, such as calibration, adjustment, and compatibility with ROTRONIC measurement transducers and hand-held meters, are also available in the AirChip4000 technology of the LDP probe.

The dewpoint measurement and its equalization time depend on many factors. ROTRONIC therefore provides a measurement chamber (see figures 2, 4, 8 and 9) with a pre-set airflow of 1 litre per minute at 8 bar pressure. In the measurement chamber, the probe experiences constant airflow, which allows very fast, precise dewpoint measurement. The equalization time of the measurement value may vary widely between applications. It depends on the flow velocity in the measurement chamber, the absolute pressure in the measurement line, the dewpoint of the surroundings, and the surface structure of the measurement line.


**1.2.1 Construction of the probe**

Figure 1: Description of the LDP probe and its components.




Figure 2: LDP probe, screwed into LDP FC measurement chamber in POM (Polyoxymethylene).

### 1.2.2 *Measurement chambers*

The measurement chamber is available in POM (polyoxymethylene) or stainless steel. It can be ordered as a base body or as a ready-assembled unit with a quick-action compressed-air coupling and fixed valve.




Figure 3: Base body of measurement chamber (LDP MC) in POM (polyoxymethylene) and stainless steel.




Figure 4: Fully assembled measurement chamber (LDP FC) in POM (polyoxymethylene) and stainless steel.

## 2 Mechanical Dimensions

### 2.1 Probe

The LDP probes are available with 2 metres or 5 metres of cable. Figures 5 and 6 show the sensor head of the low-dewpoint probe.




Figure 5: Dimensions of the LDP probe head in mm




Figure 6: 3D drawing of the probe head.

## 2.2 Measurement Chamber

The measurement chamber is available as a complete unit with fixed valve and quick-action compressed-air coupling.




Figure 7: Dimensions (in mm) of the complete measurement chamber with LDP FC fixed valve.

Customized applications can build on the base body.




Figure 8: Dimensions (in mm) of the LDP MC measurement chamber base body.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page <b>9 of 29</b>

### 3 General

#### 3.1 Power Supply

For full functionality, the LDP probe requires a power supply at 3.3 .. 5 VDC (1.5 mA).

#### 3.2 Measured Parameters

The ROTRONIC LDP probe calculates the dewpoint/frostpoint from the humidity measurements taken by the HYGROMER® LDP-1 dewpoint sensor and temperature measurements from a Pt1000 temperature sensor. The values for dewpoint, frostpoint and temperature are transmitted in the form of a digital signal.

#### 3.3 UART Interface

The LDP probe has a serial digital UART interface (UART: Universal Asynchronous Receiver Transmitter), which is compatible to the HF5, HF8, and PF4 measurement transducers, HP22 and HP23 hand-held meters, and the AC3001 service cable (probe UART to PC USB). With the HW4 software (Professional Version) installed on the PC, the following functions are available:

- Recording and display of current measured and calculated values;
- Configuration of the probe;
- Calibration and adjustment of the probe.

Connect the probe to the PC as follows:

- Connect the LDP probe to a USB port of an MS Windows PC via AC3001 (UART-to-USB adapter).
- Use HW4 to search for and include the unit:  
*Units and groups > Search for units > Search for USB master*

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page <b>10</b> of 29

### 3.4 HW4 Software Compatibility

LDP series probes are completely integrated in the HW4 software at V3.4.0 or higher.

### 3.5 Sensor Filter

The probe must always be protected from soiling by a stainless-steel sinter filter. The filter has a pore size of 50 µm (ROTRONIC Order Code: SP-S15/50).

**IMPORTANT:**

You should not touch the filter with your bare hands. Soiling of any kind can impair the precision and response time of the measurements.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page 11 of 29

## 4 User-Specific Settings and Functions

The LDP probe has a large number of functions and settings that users can adapt to suit their requirements. This chapter contains an overview of all functions and their default settings.

### 4.1 General AirChip4000 Functions

This section outlines the general functions of the AirChip family available in the LDP probes. All functions can be used with the HW4 software and are described in detail in the manual **E-M-HW4v3-LDP-001**.

#### 4.1.1 *Calibration and adjustment*

##### Dewpoint, frostpoint and temperature adjustment

- Multi-point dewpoint/frostpoint calibration and adjustment
- Single-point temperature calibration and adjustment
- Time stamp for each calibration and adjustment point
- Saving and display of the last adjustment data and values
- Creation of a calibration and adjustment log

#### IMPORTANT:

ROTRONIC recommends that you send the probe back to the manufacturer for calibration and adjustment. ROTRONIC has a professional, high-quality calibration system that guarantees optimal, certified results.

#### 4.1.2 *Write Protection for Device Settings*

##### Write protection

The device settings can be protected by a password. Password protection is configured with the HW4 software.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
Document Name	Document Type
	Page 12 of 29

#### 4.1.3 Alarm Functions

<b>Out-of-limits value alarm</b>
Specific limit values can be defined for dewpoint, frostpoint and temperature; an alarm is shown in the HW4 software when they are exceeded.
The alarm function can be enabled or disabled.
<b>Bad sensor alarm</b>
This function is a fixed component of the LDP probe firmware and cannot be disabled. In the case of a short circuit or power interruption at the dewpoint or temperature sensor, a digital alarm is triggered and displayed in the HW4 software. While the alarm is pending, the values for dewpoint/frostpoint are set to the current temperature value (e.g. at 23 °C (73 °F), the display shows 23 °C (73 °F) $T_d$ ). In the case of a temperature-sensor alarm, the temperature value is set to -99 °C (-146 °F).

#### 4.1.4 Threshold Monitoring

<b>Automatic threshold monitoring</b>
Threshold monitoring can be enabled or disabled (only with HW4 v3.5).  Enabled: (only if wished by customer this function can be enabled by factory) Threshold monitoring becomes active below a measured value of -69.9 °C (-94 °F) $T_d$ . It then sets the measured value to -66 °C (-89 °F) $T_d$ . This allows the user to observe changes in the measured values in the lower limit range.  Disabled: (default setting) The value limits integrated in the probe prevent it from measuring values below -70 °C (-94 °F) $T_d$ . The output measurement value is limited at -70 °C (-94 °F) $T_d$ .

# E-M-HC2-LDP-V1\_06

Document Code

## HC2-LDP Manual

Instruction manual

Document Type


Document Name

Page 13 of 29

There is no static system without any changes. Systems are always in motion and this must be monitored. In this case the threshold monitoring can be used when changes of the value are important even below  $-70^{\circ}\text{C}$   $T_d$ . The following graph shows the drying process of a probe after connecting it to a compressed air system. After reaching  $-70^{\circ}\text{C}$   $T_d$  the value jumps back to  $-66^{\circ}\text{C}$   $T_d$  and goes further down until the system is stable.



In this case a compressed air system is monitored which is slightly below  $-70^{\circ}\text{C}$   $T_d$ . With the threshold monitoring activated, variations (e.g.: opening of valves, extensive use of air, leakage, etc.) in the system can still be seen.



<b>E-M-HC2-LDP-V1_06</b>		
Document Code		
<b>HC2-LDP Manual</b>		<b>Instruction manual</b>
		Document Type
Document Name		Page 14 of 29

## 4.2 Factory Defaults

All factory defaults and functions of the LDP probe are described below. The HW4 software (V3.4.0 or higher) is needed to change them. An AC3001 service cable can be used to connect the probe to the PC.

Configurable Setting	Factory Default
Unit of measurement (metric/English)	Metric
Automatic threshold-value monitoring	Disabled
Dewpoint/frostpoint	Dewpoint

Functions	Factory Default
Adjustment	Three-point dewpoint adjustment Single-point dewpoint adjustment
Write protection	Disabled
Measured value alarm, invalid measured value (digital alarm)	Disabled
Simulator mode	Disabled

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page 15 of 29

## 5 Mechanical Installation

### 5.1 Choice of Measurement Point

There are several points to note when choosing the point at which to measure the dewpoint/frostpoint.

#### 5.1.1 *Positioning the probe and the measurement chamber*

The probe can be screwed directly into a line (max. 100 bar), or plugged into a compressed-air connection (max. 16 bar) with a measurement chamber (see figure 9).

**IMPORTANT:** The measurement chamber must not be connected directly to the compressed air supply without a properly screwed-in probe, and the probe must not be removed from the measurement chamber when the chamber is still under pressure.

It is recommended that you position the probe close to the application to be monitored. Long pipelines from the measurement chamber to the main line that is to be monitored make measurement unnecessarily difficult with regard to amplitude and reaction time. In order to prevent any drop in environmental humidity and pressure from affecting the measurements, there must be no leaks (except for the outlet valve of the measurement chamber).



Figure 9: POM measurement chamber LDP-FCPB1 with compressed-air connection (left), probe insertion thread (right), and fixed valve (top).

#### 5.1.2 *Positioning of probe with regard to temperature specifications*

Care must be taken that the temperature specifications of the probe are not exceeded by the medium being measured.

If the temperature is too high, the medium must be cooled down. To avoid condensation, the dewpoint must, however, be lower than the ambient temperature.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page <b>16</b> of 29

### 5.1.3 *Filter soiling*

Oil and dust particles in the gas line affect the measurement precision, reaction time and lifetime of the probe. The probe's filter cannot filter out particles less than 50 µm in diameter. In this case, it is advisable to install a more effective external filter upstream. The drop in pressure or reduction in flow volume that this causes can impair the measurement precision and reaction time. However, the measurement in its steady state would remain stable over time, even in this setup.

### 5.1.4 *Pressure at the probe head*


The maximum permanent pressure for the probe is 100 bar. Please note that the complete measurement chamber with its quick-action coupling and fixed valve is only approved for up to 16 bar.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page 17 of 29

## 5.2 Installation of the Probe

As soon as the measurement point has been determined, the probe can be installed as follows:

1. Push the sealing ring over the filter onto the G1/2" threaded shaft. Use a new sealing ring for every new installation of the probe.
2. Remove the red protective cap. Do not touch the probe filter with your bare hands.
3. First, screw the probe finger-tight into a G1/2" thread.
4. Tighten the probe with a torque wrench set to 25 Nm.  
**Important:** The probe must be tightened using only its 27 mm hex section.
5. The probe can be attached only to a transmitter, a hand-held meter or a PC.
  - Connection to a PC: The plug of the probe is connected to a PC via the AC3001 service cable.
  - Connection to a transmitter (HF5, HF8 or PF4): The plug of the probe is attached directly to the probe input on the transmitter.
  - Connection to a hand-held meter (HP22, HP23): The plug of the probe is attached directly to the probe input on the hand-held meter.



### 5.3 Pin Configuration of Plug Connector

The low-dewpoint probe has a ROTRONIC E2 plug that is compatible with measurement transducers HF5, HF8, PF4, HP22 and HP23, and with the AC3001 service cable. Its pin configuration is shown in figure 10.




Figure 10: Pin configuration of the E2 plug

Pin	Name	Function
1	V (+)	3.3 .. 5 VDC
2	GND	Supply ground
3	RxD	UART reception line
4	TxD	UART transmission line
5	-	Not connected
6	-	Not connected
7	-	Not connected

<b>E-M-HC2-LDP-V1_06</b>		
Document Code		
<b>HC2-LDP Manual</b>		<b>Instruction manual</b>
		Document Type
Document Name		Page <b>19</b> of 29

## 6 Maintenance

This chapter describes calibration, adjustment, validation and filter changes as well as the tools necessary for this, e.g. service cable.

### 6.1 Service Cable

An **AC3001** service cable can be used for maintenance and calibration. The service cable has an internal power supply of 3.3 VDC. Alternatively, the connection to the PC can be made via an **HF5**, **HF8** or **PF4** transmitter, or an **HP22/HP23** hand-held meter.

### 6.2 Calibration

Both the **Pt1000-RTD temperature sensor** used in the probe and the corresponding electronics are very stable and **do not normally need to be calibrated** after initial factory calibration. The long-term stability of the **HYGROMER® LDP-1 sensor** from ROTRONIC is normally better than  $1.^\circ\text{C T}_d$  per year. For maximum accuracy, the calibration of the probe should be checked every **12 months**. Applications in which the probe is exposed to considerable contamination could require more frequent checks.

#### **IMPORTANT:**

ROTRONIC recommends that you send the probe back to the manufacturer for calibration and adjustment. ROTRONIC has a professional, high-quality calibration system that guarantees optimal, certified results.

Customers can carry out the adjustments themselves, using the HW4 software.

- Start the HW4 software and look for the connected LDP probe (*HW4 Menu > Devices and Groups > Search for Devices > Search for USB Master*)
- The device tree of the LDP probe lists all calibration and adjustment functions.
- **At least 2 calibration points** must be recorded, if adjustments are to be made at a later date.
- The adjustment points must be recorded in the sequence from dry to moist; the lowest calibration point must be at the lower limit of the measurement range, with two further points in the range  $-10 \dots 10^\circ\text{C}$  ( $14 \dots 50^\circ\text{F}$ )  $T_d$ .
- You will find further instructions in the LDP sensor software manual, **E-M-HW4v3-LDP-001**.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page <b>20</b> of 29

### 6.2.1 *Dewpoint and temperature adjustment*

The probe should be calibrated at least once a year, and adjusted if necessary. ROTRONIC strongly advises that you return the probe to the factory for this. If adjustment on site is necessary, the probe should be readjusted at the following dewpoint values.

- -60 °C T<sub>d</sub> ( $\pm 5$  °C T<sub>d</sub>)
- -10 °C T<sub>d</sub> ( $\pm 5$  °C T<sub>d</sub>)

You can add further points as necessary. It is advisable to adjust the probe in the order from dry dewpoint value to humid dewpoint value. Ensure that the probe is in steady state at each dewpoint, i.e. the measured dewpoint value no longer changes.

The equipment needed for adjustment (dewpoint reference, low dewpoint generator) has a decisive influence on the measuring accuracy of the probe.

Calibration/adjustment can be performed with the HW4 software. Should it be necessary to adjust the LDP probe, this can be done with the HW4 software. You will find further instructions in the HW4 software manual **E-M-HW4v3-LDP-001** for the LDP probe.

### 6.3 **Probe**

It is recommended to carry out maintenance on the probe at least once a year (or, in the case of measurements in oil-polluted air, every 3 months). To do this, remove the filter and submit it to a visual inspection. Soiled parts, such as measurement chamber, fixed valve or filter, must be cleaned in an ultrasonic bath and alcohol, or exchanged. If the sensor is soiled, the probe must be returned to the manufacturer.

#### **IMPORTANT:**

The sensor filter must be removed very carefully, because touching the trace moisture sensor can destroy it! The filter should not be touched with bare hands. Any improper maintenance can lead to a reduction in the precision of measurements.

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page <b>21</b> of 29

## 7 Firmware Update

The LD Probe firmware can be updated with the HW4 software. Firmware updates are available for downloading on the ROTRONIC website. Updating the firmware requires that the probe be connected via an AC3001 service cable to a PC with HW4 software installed on it.

### **IMPORTANT:**

There must be a permanent connection to the computer and a stable power supply during the update process.

### **Procedure:**

- Connect the probe to the service cable, or to an HF5, HF8, PF4, HP22 or HP23.

### **Remark:**

As of 2014, the service cables work with a standard USB driver. Older service cables need a ROTRONIC USB driver, which is installed on the PC together with the HW4 software. For detailed information on the ROTRONIC USB driver, see the manual **E-M-HW4v3-Main**.

- Download the latest LDP probe firmware to your PC from the ROTRONIC website.
- Launch the HW4 software on the PC and search for the LDP probe.

*HW4 Main Menu > Devices and Groups > Search for Devices > Search for USB Master / Search for Ethernet Master*

- Select the Update option under Device Manager in the menu Extras > Firmware Update

For further information, see the manual **E-M-HW4v3-LDP-001**.

# E-M-HC2-LDP-V1\_06

Document Code

## HC2-LDP Manual

Instruction manual

Document Type

Document Name

Page 22 of 29

## 8 Technical Data

General	
Device type	Low-dewpoint/temperature probe
Output signal type	Digital
Compatible	HF5, HF8, PF4, HP22, HP23, AC3001, HW4 (V3.4.0 or higher)

Power supply	
(V+)	3.3...5 VDC
Current consumption	<1.5 mA
Electrical connection	E2 plug (7-pin)
Polarity protection	Mechanical

Dewpoint / frostpoint measurement	
Sensor	HYGROMER® LDP-1
Measurement range	-70...85 °C (-94...185 °F) $T_d$
Measurement precision at 23 °C (73 °F)	±2 K $T_d$ (-50...20 °C / -58...68 °F $T_d$ ) ±3.5 K $T_d$ (-60...-50 °C / -76...-58 °F $T_d$ )
Long-term stability	<1 °C $T_d$ /year
Response time T63: at -50 °C (-58 °F) $T_d$	Typically: 10 mins. (humid to dry) Typically: 13 s (dry to humid)
Calibration / adjustment	Multi-point calibration / adjustment

Temperature measurement	
Sensor	Pt1000, Class B 1/3 DIN
Measurement range	-40...85 °C / -40...185 °F
Measurement precision at 23 °C (73 °F)	±0.2 K
Response time T63: at -50 °C (-58 °F) $T_d$	Typically: 15 mins.
Calibration / adjustment	Single-point calibration / adjustment

Startup time and measurement interval	
Startup time	4 s (typical)
Measurement interval	2 s (typical)

# E-M-HC2-LDP-V1\_06

Document Code

## HC2-LDP Manual

Instruction manual

Document Type

Document Name

Page 23 of 29

### Digital interface

Interface type	Serial (UART) Baud rate : 19,200 Parity : none Data bits : 8 Stop bits : 1
Maximum length of cable	5 m (16.4 ft)

### General specifications

Material of probe housing	Stainless steel, 1.4301
Filter material	Stainless-steel sinter filter, pore size 50 µm
IP protection	IP65
Pressure resistance, sensor component	16 bar (100 bar with LDP-MCS measurement chamber)
Dimensions	145 x 31 mm (probe head) Hex nut: 27 mm
Weight	260 g (9.2 oz)

### Conformity

CE / EMC immunity	EMC Directive: until 20.04.2016: 2004/108/EG from 20.04.2016: 2014/30/EU  EN 61000-6-1: 2007, EN 61000-6-2: 2005 EN 61000-6-3: 2007+A1:2011+AC:2012, EN 61000-6-4: 2007+A1:2011  Performance criterion: <a href="http://www.rotronic.com">www.rotronic.com</a>
Solder type	Lead free (RoHS directive 2011/65/EU)
FDA & GAMP directives	Conforms

### Environmental limits

Storage and transit	-40 .. 85 °C (-40 .. 185 °F) non-condensing
Operation	-40 .. 85 °C (-40 .. 185 °F) non-condensing
Critical environments	Trace moisture sensor: per DV04-14.0803.02 – critical chemicals

<b>E-M-HC2-LDP-V1_06</b>		
Document Code		
<b>HC2-LDP Manual</b>		<b>Instruction manual</b>
		Document Type
Document Name	Page <b>24</b> of 29	

## 8.1 Chemical compatibility of LDP-1 sensor

Name	Formula	Concentration	Duration of exposure	Deviation
Carbon dioxide	CO <sub>2</sub>	10,000 ppm	20 min	±0.2 °C T <sub>d</sub>
Carbon monoxide	CO	100 ppm	20 min	±0.3 °C T <sub>d</sub>
Ammonia	NH <sub>3</sub>	50 ppm	20 min	±0.3 °C T <sub>d</sub>
Nitrogen dioxide	NO <sub>2</sub>	5 ppm	20 min	±0.1 °C T <sub>d</sub>
Hydrogen	H <sub>2</sub>	3,500 ppm	20 min	±0.2 °C T <sub>d</sub>
Hydrogen sulfide	H <sub>2</sub> S	1 ppm	20 min	±0.6 °C T <sub>d</sub>
Ethylene oxide	C <sub>2</sub> H <sub>4</sub> O	2...3 %	18 cycle a 90 min	±4 °C T <sub>d</sub>
Methane	CH <sub>4</sub>	100 %	11 days	±0.6 °C T <sub>d</sub>
Motor oil	-	100 %	Long-time exposure	±0.4 °C T <sub>d</sub>
Acetone	CH <sub>3</sub> COCH <sub>3</sub>	100 %	1 h	±0.4 °C T <sub>d</sub>
Toluol	C <sub>6</sub> H <sub>5</sub> CH <sub>3</sub>	100 %	3 h	±0.8 °C T <sub>d</sub>
Xylene	C <sub>6</sub> H <sub>5</sub> (CH <sub>3</sub> ) <sub>2</sub>	100 %	3 h	±0.2 °C T <sub>d</sub>
Helium	He	10 %	20 h	±0.2 °C T <sub>d</sub>
Hydrogen peroxide	H <sub>2</sub> O <sub>2</sub>	800 ppm	38 days	±1.2 °C T <sub>d</sub>

**Note:**

The deviations were measured at 23°C and 80 %RH (19.37 °C T<sub>d</sub>).

## E-M-HC2-LDP-V1\_06

Document Code

### HC2-LDP Manual

Instruction manual

Document Type

Document Name

Page 25 of 29

## 8.2 Technical data, measurement chamber

Base body (LDP-MC)	
Ambient temperature	-40...85 °C (LDP-MCP, POM) -50...100 °C (LDP-MCS, stainless steel)
Pressure resistance	16 bar (LDP-MCP, POM) 100 bar (LDP-MCS, stainless steel)
Dimensions	40 x 106 mm
Connections	G1/2" thread for LDP probe 2 x G1/4" thread
Housing material	POM (LDP-MCP) Stainless steel 1.4301 (LDP-MCS)

### Measurement chamber with quick-action coupling and fixed valve (LDP-FC)

Air flow	1 litre per min. at 8 bar
Measurement range	-40 .. 85 °C (POM version) -50 .. 100 °C (stainless-steel version)
Pressure resistance	16 bar
Dimensions	40 x 106 mm
Connections	G1/2" thread for LDP probe Compressed-air nipple NW7.2
Housing material	Base body: POM / stainless steel 1.4301 Valve: Zirconium oxide Quick-action coupling: Brass

## 9 Accessories

All accessories for ROTRONIC products are to be found in the manual **E-M-HC2-accessories**. The main service cables for the LDP probe are also listed there:

Order Code	Description	
LDP-MCP	Measuring chamber base body, POM	
LDP-MCS	Measuring chamber base body, stainless steel	
LDP-FCPB1	Measurement chamber with fixed coupling and quick-action compressed-air coupling, POM	
LDP-FCSB1	Measurement chamber with fixed valve and quick-action compressed-air coupling, stainless steel	
AC4001-B	Replacement fixed valve	

# E-M-HC2-LDP-V1\_06

Document Code

## HC2-LDP Manual

Instruction manual

Document Type

Document Name

Page 27 of 29

AC3001	USB/UART service cable with E2 socket	
SP-S15/50	Stainless-steel sinter filter, 50 µm	
AC4003	Sealing ring	

<b>E-M-HC2-LDP-V1_06</b>	
Document Code	
<b>HC2-LDP Manual</b>	<b>Instruction manual</b>
	Document Type
Document Name	Page <b>28</b> of 29

## 10 Additional Documents

Document Name	Contents
<b>E-M-HW4v3-LDP-001</b>	HW4 software manual for operation of the LDP probe.
<b>E-M-HW4v3-Main</b>	HW4 software version 3: general instructions and description of functions.
<b>E-M-HW4v3-DIR</b>	List of all HW4 manuals.
<b>E-M-HC2-accessories</b>	Accessories such as service cables, calibration accessories, filters for probes, transmitters.

**Remark:**

The document titles are supplemented by the appended version numbers. These version numbers have been omitted in the above table.

<b>E-M-HC2-LDP-V1_06</b>		
Document Code		
<b>HC2-LDP Manual</b>		<b>Instruction manual</b>
		Document Type
Document Name		Page <b>29</b> of 29

## 11 Document Approval

<b>Version</b>	<b>Date</b>	<b>Remark</b>
V1_00	November 2014	Release document
V1_01	November 2014	Graphic of threshold monitoring added
V1_02	March 2015	Response time dry to humid added
V1_03	March 2015	Threshold monitoring is disabled by factory
V1_04	September 2015	Calibration / adjustment specification included
V1_05	November 2015	Chemical compatibility of LDP-1 sensor added
V1_06	May 2016	Update CE guideline